



Certificate of Analysis

Abbott Ireland Diagnostics Division,
Lisnamuck,
Longford,
Co. Longford,
Ireland

Product Name: Microalbumin Calibrators
List number: 08P0404
Lot number: 74174UD00

SCP / revision 8P040004/005

Expiry Date: 2026-12-03

| Component name | List Number | Lot Number | Production date | Expiry date |
|----------------|-------------|------------|-----------------|-------------|
| Microalb Cal 1 | 08P04B | 74169UD00 | 2024-12-03 | 2026-12-03 |
| Microalb Cal 2 | 08P04C | 74170UD00 | 2024-12-03 | 2026-12-03 |
| Microalb Cal 3 | 08P04D | 74171UD00 | 2024-12-03 | 2026-12-03 |
| Microalb Cal 4 | 08P04E | 74172UD00 | 2024-12-03 | 2026-12-03 |
| Microalb Cal 5 | 08P04F | 74173UD00 | 2024-12-03 | 2026-12-03 |

Acceptance Table: Grand Mean Concentration (ug/mL)

| Description | Lower Limit | Upper Limit | Value | Result |
|-----------------|-------------|-------------|-------|------------|
| REF Control L 1 | 27.00 | 33.00 | 29.35 | Acceptable |
| REF Control N 2 | 81.90 | 98.10 | 89.93 | Acceptable |
| REF Control 3 | 13.4 | 17.4 | 14.6 | Acceptable |
| REF Control 4 | 143.7 | 177.3 | 156.7 | Acceptable |

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

DAVOR DZANAK Du Dzak
Printed name and signature

Date:

18 Mar 25

